

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
-------------	--------	-----------------------------	--------------------------------------	---------	--------------	---------

Illumina Short-Read Sequencing

<i>DNA Sequencing</i>	General remarks and requirements for DNA samples: <ul style="list-style-type: none"> - Buffer: TRIS-HCL 10 mM or lowTE - 260/280 ratio 1.8-2.0 (or according to quality column) - In case of genome sequencing: high molecular for best results <p>All samples of an order must be adjusted to a uniform concentration within the specifications. Samples that do not meet the requirements listed here cannot be processed and will be rejected!</p>					
<i>Whole Genome - Low input (incl. PCR)</i>	150 – 300ng	≥ 15	> 10	see general remarks above	-	FFPE possible with constraints in output quality
<i>Whole Genome - PCR Free</i>	> 300 ng	≥ 15	> 20	see general remarks above	-	
<i>Bacterial genome sequencing</i>	15 ng	≥ 15	1	see general remarks above		
<i>Whole Exome Sequencing</i>	> 75 ng	≥ 15	5	see general remarks above	-	FFPE possible if amount based on qPCR measurement is sufficient
<i>Gene Panel</i>	600 ng	≥ 15	40	260/280 ratio 1.7-2.2 260/230 ratio 1.2	-	-
<i>Amplicons</i>	> 75 ng	≥ 15	5	260/280 ratio > 1.5	-	less material possible on request
<i>Microbiome 16S</i>	150 ng	≥ 15	10	260/280 ratio > 1.5	-	less material possible on request
<i>Microbiome Shotgun Metagenomics</i>	75 ng	≥ 15	5	260/280 ratio > 1.5	-	less material possible on request

	Application	Amount	Volume (μL)	Conc. (ng/ μL)	Quality	Size (bp)	Comment
Illumina Short-Read Sequencing							
<i>RNA Sequencing</i>							
General remarks and requirements for totalRNA samples:							
<ul style="list-style-type: none"> - Buffer: nuclease free water - DNase treated and cleaned up - 260/280 ratio >2.0 - RQN >=8; RQN Δ between samples <1 - ultra low input (< 1 ng total amount) possible, but must be planned beforehand 							
All samples of an order must be adjusted to a uniform concentration within the specifications.							
Samples that do not meet the requirements listed here cannot be processed and will be rejected!							
	<i>Whole Transcriptome – standard input</i>	>375 ng	>=15	25 - 80	see general remarks above	-	-
	<i>Whole Transcriptome - low input</i>	>15 ng	>=15	1 - 5	see general remarks above	-	very low input amount may lead to constraints in output quality
	<i>Expression Profiling (mRNA) – standard input</i>	>375 ng	>=15	25 - 80	see general remarks above	-	-
	<i>Expression Profiling (mRNA) – low input</i>	>15 ng	>=15	1 - 5	see general remarks above	-	very low input amount may lead to constraints in output quality
	<i>3'Prime RNA Seq -standard input</i>	>375 ng	>=15	25 - 80	see general remarks above	-	-
	<i>3'Prime RNA Seq -low input</i>	>15 ng	>=15	1 - 5	see general remarks above	-	very low input amount may lead to constraints in output quality
	<i>Small RNA Profiling (miRNA, Inc-RNA etc.)</i>	50 ng	>=15	50	see general remarks above	-	Low input possible on request

Application	Amount	Volume (μL)	Conc. (ng/ μL)	Quality	Size (bp)	Comment
<i>Sequencing of prepared NGS libraries</i>	50 ng	≥ 15	5	No primer dimer residuals	-	Buffer: TRIS-HCL 10 mM
10X Single-Cell RNA-Sequencing						
<i>Single-Cell Transcriptomics</i>	1.000 – 30.000 cells /sample	50 μl	500-1.000 cells/ μl	Vitality >70 %	-	-
<i>Spatial Transcriptomics</i>	Tissue sections placed on Visium slides, RNA extracted from tissue section RIN >7					
Bionano Optical Mapping						
<i>Bionano Saphyr Chip</i>	1 μg	-	>36	260/280 = 1.8 260/230 = 2.0-2.2	Megabase range	-

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
Pacific Biosciences Long-Read Sequencing						
Revio System						
<i>Transcriptome</i>						
	<i>Iso-Seq</i>	>2-3 μg	>15	>200	RIN 8-10 260/280 = 2.0 260/230 = 2.2	- DNase digested, buffer: RNase-free water
<i>Whole Genome Sequencing</i>						
	<i>HiFi Reads – Standard Protocol</i>	$\geq 8 \mu\text{g}$	>120	~ 70	260/280 = 1.8 260/230 = 2.0-2.2	>50kb homogenous HMW DNA, RNase digested, buffer: EB, DNA size distribution: 80% $\geq 50\text{Kb}$
	<i>HiFi Reads – Low Input</i>	>3 μg	>50	60	260/280 = 1.8 260/230 = 2.0-2.2	>50kb homogenous HMW DNA, RNase digested, buffer: EB, DNA size distribution: 80% $\geq 50\text{Kb}$
	<i>HiFi Reads – Ultra-Low Input</i>	>50 ng	>15	>3	260/280 = 1.8 260/230 = 2.0-2.2	>50kb homogenous HMW DNA, RNase digested, buffer: EB DNA size distribution: 80% $\geq 50\text{Kb}$
<i>Trageted Sequencing</i>						
	<i>Amplicon Sequencing</i>	2-4 μg (depends on size)	50	-	-	- clean, target-specific, buffer: EB
<i>Metagenomics</i>						
	<i>Multiplexed Microbial</i>	>2-3 μg per sample	20-100	-	260/280 = 1.8 260/230 = 2.0-2.2	>30kb Homogenous HMW DNA, RNase digested, buffer: EB DNA size distribution: 80% $\geq 30\text{Kb}$

Application	Amount	Volume (μ L)	Conc. (ng/ μ L)	Quality	Size (bp)	Comment
<i>Kinnex workflows</i>						
<i>Full-length RNA</i>	2-3 μ g (depends on # samples to concatenate)	50	> 200	RIN 8-10 260/280 = 2.0 260/230 = 2.2	-	DNase digested, buffer: RNase-free water up to 8-fold concatenation
<i>16S rRNA Amplicons</i>	>500 ng - 3 μ g (depends on size)	50	-	-	~ 1500 bp	Purified amplicons up to 12-fold concatenation (~ 19 Kb) Applicant must order Kinnex 16S Forward and reverse primers. From any oligo vendor, HPCL purification is recommended.
<i>Single-cell RNA</i>	50 ng of 10x Chromium 3` or 5` single cell DNA.	25	-		500-1500 bp	Purified up to 16-fold concatenation Single-cell cDNA generated using the 10x Chromium Next GEM Single Cell 3' kit v3.1/ 5' kit v2 standard throughput. cDNA in the respective buffer.

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
Sequel II/IIe Systems						
<i>Transcriptome (Iso-Seq)</i>	>2 μg	>15	>130	RIN 8-10 260/280 = 2.0 260/230 = 2.2	-	DNase digested, buffer: RNase free water
<i>Whole Genome Sequencing</i>						
<i>HiFi Reads – Standard Protocol</i>	>6 μg	>120	50	260/280 = 1.8 260/230 = 2.0-2.2	>50kb	homogenous HMW DNA, RNase digested, buffer: EB DNA size distribution: 80% \geq 50Kb
<i>HiFi Reads – Low Input</i>	>1 μg	>50	20	260/280 = 1.8 260/230 = 2.0-2.2	>50kb	homogenous HMW DNA, RNase digested, buffer: EB DNA size distribution: 80% \geq 50Kb
<i>HiFi Reads – Ultra-Low Input</i>	>30 ng	>15	2	-	>50kb	homogenous HMW DNA, RNase digested, buffer: EB DNA size distribution: 80% \geq 50Kb
<i>Continous Long Reads</i>	>6 μg	>120	50	260/280 = 1.8 260/230 = 2.0-2.2	>50kb	homogenous HMW DNA, RNase digested, buffer: EB
<i>Amplicon Sequencing</i>	500 ng – 3 μg (depends on size)	50	-	-	-	clean, target-specific, buffer: EB
<i>Multiplexed Microbial</i>	>1 μg per sample	20-100	-	260/280 = 1.8 260/230 = 2.0-2.2	>30kb	-

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
Oxford Nanopore Long-Read Sequencing						
<i>Transcriptome</i>						
<i>Direct mRNA Sequencing</i>	>500 ng polyA+ or 1.5 μg Total-RNA	>12	>40	260/280 = 2.0 260/230 = 2.0-2.2	-	DNase digested, buffer: RNase free water
<i>cDNA Sequencing</i>	>200 ng polyA+ or >1 μg Total-RNA	>10	>20	260/280 = 2.0 260/230 = 2.0-2.2	-	200 ng cDNA can be used as input DNase digested, buffer: RNase free water
<i>cDNA PCR Sequencing</i>	>20 ng polyA+ or >600 ng Total-RNA	>15	>1.3	260/280 = 2.0 260/230 = 2.0-2.2	-	DNase digested, buffer: RNase free water
<i>Whole Genome Sequencing</i>						
<i>Ligation Sequencing</i>	>2 μg	>50	>40	260/280 = 1.8 260/230 = 2.0-2.2	>30 kb	homogenous HMW DNA, RNase digested, buffer: EB
<i>Ligation Sequencing with Size-Selection</i>	>10 μg	>50	>200	260/280 = 1.8 260/230 = 2.0-2.2	>30 kb	homogenous HMW DNA, RNase digested, buffer: EB
<i>Ligation Sequencing Whole-Genome Amplification</i>	>1 ng	>10	>0.1	260/280 = 1.8 260/230 = 2.0-2.2	>30 kb	homogenous HMW DNA, RNase digested, buffer: EB
<i>Native Barcoding</i>	>1 μg per sample	>50	>20	260/280 = 1.8 260/230 = 2.0-2.2	>30 kb	homogenous HMW DNA, RNase digested, buffer: EB
<i>Ultra Long Reads</i>	>50 μg	>750	>70	260/280 = 1.8 260/230 = 2.0-2.2	>100 kb	homogenous UHMW DNA, RNase digested, buffer: EEB (please inquire about EEB buffer)
<i>Rapid - gDNA</i>	>300 ng	>10	>30	260/280 = 1.8 260/230 = 2.0-2.2	>30 kb	homogenous HMW DNA, RNase digested, buffer: EB

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
-------------	--------	-----------------------------	--------------------------------------	---------	--------------	---------

Oxford Nanopore Long-Read Sequencing

<i>Amplicon Sequencing</i>						
	<i>Amplicons by Ligation</i>	>300 ng per sample	>20	>15	260/280 = 1.8 260/230 = 2.0-2.2	- clean, target-specific, buffer: EB
	<i>Rapid - 16S</i>	>20 ng	>20	>1	260/280 = 1.8 260/230 = 2.0-2.2	- clean, target-specific, buffer: EB

Sanger Sequencing

<i>Full Service</i>						
	<i>Plasmid</i>	>500 ng	-	100 – 250	260/280 = 1.8 260/230 = 2.0-2.2	<5 kb Buffer: 10 mM Tris/HCl or water Primer concentration 10 μM
		>800 ng	-	150 - 600	260/280 = 1.8 260/230 = 2.0-2.2	5 kb – 15 kb Buffer: 10 mM Tris/HCl or water Primer concentration 10 μM
		>1,5 μg	-	>600	260/280 = 1.8 260/230 = 2.0-2.2	>15 kb Buffer: 10 mM Tris/HCl or water Primer concentration 10 μM
<i>Full Service / Xpress Service</i>						
	<i>PCR-product</i>	max. 10 ng	-	<2,5	260/280 = 1.8 260/230 = 2.0-2.2	<100 bp purified PCR products Primer concentration 10 μM
		50 ng	-	< 5ng	260/280 = 1.8 260/230 = 2.0-2.2	100 bp – 1 kb purified PCR products Primer concentration 10 μM

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
	400 ng	-	20 - 50	260/280 = 1.8 260/230 = 2.0-2.2	>1 kb	purified PCR products Primer concentration 10 μM

Sanger Sequencing

<i>Xpert Service</i>		-	-	-	-	Sample requirements as for Full Service
<i>Pre-mixed Service</i>						
	<i>Plasmid</i>	300 – 600 ng	-	-	-	<10 kb 5 pmol Primer/reaction 7,5 μL total volume
		>700 ng	-	-	-	>10 kb 5 pmol Primer/reaction 7,5 μL total volume
	<i>PCR-products</i>	<100 ng	-	-	-	100 bp – 1 kb 5 pmol Primer/reaction 7,5 μL total volume
		>100 ng	-	-	-	>1 kb 5 pmol Primer/reaction 7,5 μL total volume
<i>Ready-to-load Service</i>						
	<i>Plasmid, PCR-products</i>	-	20	-	-	Purified sequencing reaction

Fragment Analysis

	<i>Ready-to-run</i>	-	20	-	-	<600 bp Fully prepared samples dissolved in formamide
	<i>Pre-pared</i>	-	10	-	-	<600 bp samples dissolved in formamide without length standard

Application	Amount	Volume (μL)	Conc. ($\text{ng}/\mu\text{L}$)	Quality	Size (bp)	Comment
<i>STR analysis</i>	50 ng	-	5	260/280 = 1.8 260/230 = 2.0-2.2	-	Buffer: 10 mM Tris/HCl, low TE or water